



400 E. Lugbill Road Archbold, OH 43502  
Phone 419.445.6060 Fax: 419. 446.2631 Customer Service 800.537.2521 www.gendroninc.com

To: Steve Cotter, President/CEO  
From: Nick Stanfield, Designer  
Date: 2/7/07  
Subject: 1000MR and 4000MR Magnetic Field Interaction Testing

Tests were conducted on Gendron's stretcher and wheelchair labeled as "Non-Magnetic" to determine the product's safety in a high magnetic field environment.

Tests were conducted by Shellock R & D Services, Inc. on 10/30/06. Each product was evaluated with a powerful hand-held magnet prior to exposure to the MRI environment to determine the presence of ferromagnetism in the metallic components.

Once the products were deemed safe for use in an MRI environment by use of the hand-held magnet the products were moved near the MR system. Each product was moved to the bore of the MR system and rotated 360 degrees. It was observed that there were no free movements or evidence of attraction, rotation or alignment when exposed to the MR system.

Upon completion of testing it has been determined that the products may be properly labeled as "NONMAGNETIC at 3-Tesla or less." Each product is MR Conditional and can be used in the MRI environment according to the following conditions:

- Static magnetic field of 3-Tesla or less
- Spatial gradient magnetic field of 720-Gauss/cm or less

Note that the products are intended for use inside of the MRI environment (in the MR system room, close to the scanner), but are not to be used directly inside of the MR system (inside the bore of the scanner).

Please let me know if you need additional information.